

### Continuous Radon Monitor

Model Number: 1028  
Calibration Date: 03/20/2009  
Monitor Time: 10/12/2009 15:15

Serial Number: 49745162  
CF: 2.69

### Inspector Information

Radon Services of Kansas  
4932 Kimberly Ln.  
Wichita  
KS-67204  
Phone Number: 316-821-9611

### Customer Information

SAMPLE  
123 n.MAIN  
WICHITA  
KS-00000  
Phone: 000-000-0000


### Site & Condition

Wind: NA  
Year Built: NA  
SqFt: NA  
Atmospheric Condition: NA  
Structure Type: NA  
Monitor Location: NA

### Test Summary

Start Time: 09/25/2009 12:37  
End Time: 09/27/2009 12:37  
Measurement Interval(hr): 1.0  
Elapse Time: 2 Days 0 hrs  
Overall Avg: 0.5 pCi/l  
EPA Avg: 0.4 pCi/l

### Continuous Radon Monitor



b ◆ Radon pCi/l    b ● Motion Errors    e — Temp F    e — Rel % Humidity    e — Press inHg

\*\*\*\*\* 09/25/2009 \*\*\*\*\*

| Time  | Counts<br>pCi/l | Flags |
|-------|-----------------|-------|
| 13:37 | 0.4             |       |
| 14:37 | 0.4             |       |
| 15:37 | 1.5             |       |
| 16:37 | 1.9             |       |
| 17:37 | 0.4             |       |
| 18:37 | 0.4             |       |
| 19:37 | 0.0             |       |
| 20:37 | 0.4             |       |
| 21:37 | 0.0             |       |
| 22:37 | 0.7             |       |
| 23:37 | 0.4             |       |

\*\*\*\*\* 09/26/2009 \*\*\*\*\*

| Time  | Counts<br>pCi/l | Flags |
|-------|-----------------|-------|
| 00:37 | 0.0             |       |
| 01:37 | 0.0             |       |
| 02:37 | 0.7             |       |
| 03:37 | 0.0             |       |
| 04:37 | 0.0             |       |
| 05:37 | 0.7             |       |
| 06:37 | 0.0             |       |
| 07:37 | 0.4             |       |
| 08:37 | 1.1             |       |
| 09:37 | 0.7             |       |
| 10:37 | 0.4             |       |
| 11:37 | 0.4             |       |
| 12:37 | 0.0             |       |
| 13:37 | 0.4             |       |
| 14:37 | 0.0             | M     |
| 15:37 | 0.0             |       |
| 16:37 | 0.0             |       |
| 17:37 | 0.7             |       |
| 18:37 | 0.4             |       |
| 19:37 | 0.0             |       |
| 20:37 | 0.0             |       |
| 21:37 | 0.4             |       |
| 22:37 | 0.0             |       |
| 23:37 | 0.0             |       |

\*\*\*\*\* 09/27/2009 \*\*\*\*\*

| Time  | Counts<br>pCi/l | Flags |
|-------|-----------------|-------|
| 00:37 | 1.5             |       |
| 01:37 | 0.4             |       |
| 02:37 | 0.4             |       |

\*\*\*\*\* 09/27/2009 \*\*\*\*\*

| Time  | Counts<br>pCi/l | Flags |
|-------|-----------------|-------|
| 03:37 | 0.4             |       |
| 04:37 | 1.5             |       |
| 05:37 | 1.1             |       |
| 06:37 | 0.7             |       |
| 07:37 | 0.0             |       |
| 08:37 | 0.4             |       |
| 09:37 | 1.9             |       |
| 10:37 | 0.7             |       |
| 11:37 | 0.7             |       |
| 12:37 | 1.5             |       |

Error Flags:

M Motion:

Inspector Signature \_\_\_\_\_

PC Software Version: 2.1.3.0

Embedded Software Version: 106